


Ordering Information

Balloon length (mm)	Balloon Diameter (mm)						
	1.50	2.00	2.50	2.75	3.00	3.50	4.00
10	PWR-1510	PWR-2010	PWR-2510	-	PWR-3010	PWR-3510	PWR-4010
15	PWR-1515	PWR-2015	PWR-2515	-	PWR-3015	PWR-3515	PWR-4015
20	PWR-1520	PWR-2020	PWR-2520	PWR-2720	PWR-3020	PWR-3520	PWR-4020
25	PWR-1525	PWR-2025	PWR-2525	-	PWR-3025	PWR-3525	PWR-4025
30	PWR-1530	PWR-2030	PWR-2530	-	PWR-3030	PWR-3530	PWR-4030

Compliance Table

Pressure (atm)	Balloon Size (mm)						
	1.50	2.00	2.50	2.75	3.00	3.50	4.00
5	1.47	1.97	2.47	2.72	2.97	3.47	3.97
6 Nominal	1.50	2.00	2.50	2.75	3.00	3.50	4.00
7	1.53	2.03	2.53	2.78	3.03	3.53	4.03
8	1.56	2.06	2.56	2.81	3.06	3.56	4.06
9	1.59	2.09	2.59	2.84	3.09	3.59	4.09
10	1.62	2.12	2.62	2.87	3.12	3.62	4.12
11	1.65	2.15	2.65	2.90	3.15	3.65	4.15
12	1.68	2.18	2.68	2.93	3.18	3.68	4.18
13	1.71	2.21	2.71	2.96	3.21	3.71	4.21
14 Rated Burst	1.74	2.24	2.74	2.99	3.24	3.74	4.24



- A. Flexible tip section with smooth transition and short bonding to increase flexibility gradually.
- B. Semi-compliant Polyamide Elastomer balloon material combined with Slip-X™ hydrophilic coating for enhanced crossability.
- C. Power stylet transition in the middle shaft to provide a smoother force transmission from the proximal to the distal shaft.
- D. PTFE coated 2.0F hypotube for improved pushability.

POWERLINE™
PTCA CATHETER

Take the fast track to superior patient outcomes

Powerline™ PTCA catheter is CE approved.

CAUTION: The law restricts these devices to sale by or on the order of a physician. Indications, contraindications, warnings and instructions for use can be found in the product labeling supplied with each device.

Powerline, BioMatrix, Biolimus A9, Slip-X and Improving Lives Through Medical Innovation are trademarks of Biosensors International Group, Ltd.

All cited trademarks are the property of their respective owners.

Not available for sale in the United States and certain other countries
© 2014 Biosensors International Group, Ltd. All rights reserved

www.biosensors.com